

# Sound Level Meter, Class 2

## GNV-102



### Features:

- Meets IEC 60942 Class 2 requirements
- Capable of calibrating 1/2 and 1 inch microphones
- Calibrates at 94 dB and 114 dB

### Applications:

- OSHA compliance
- Noise ordinance checks
- General purpose noise surveys
- Commercial construction
- R&D Labs
- Education

### Description:

The Global Specialties GNV-102 Sound Level Calibrator (class 2) is a portable sound level calibrator to maintain the accuracy of your Sound Level Meter devices. With both a 1/2 inch and 1 inch port, calibration can be done on our GNV-100 and GNV-101 sound level meters and many more. A 40 hour battery life will ensure that your sound level meters are always properly calibrated.



## Specifications:

Operating Limits and Functions	
Measuring Level Ranges	94 dB at 20 uPa
	114 dB at 20 uPa
Output Frequency	1000 Hz $\pm$ 2 %
Accuracy	$\pm$ 0.5 dB
Total Harmonic Distortion	<3%
Reference Conditions	
Temperature	23 °C (74 °F)
Relative Humidity	50 %
Atmospheric Pressure	1013 hPa
Influence of Ambient Conditions	
Temperature Coefficient	0.005 dB / °C
Humidity Coefficient	0.005 dB / % Relative humidity
Environmental Specifications	
Operating Temperature	0 to 40 °C (32 to 104 °F)
Operating Humidity	10 to 90% Relative humidity
Storage Temperature	-10 to 50 °C (14 to 122 °F)
Storage Humidity	0 to 70% Relative humidity
Pressure	650 - 1080 hPa
Mechanical and Electrical Specifications	
Power	One 9V battery: 006P, IEC 6F22, NEDA 1604
Battery Life	Approx. 40 hours
Dimensions	113 x 63 x 44 mm (4.4 x 2.4 x 1.7")
Weight	Approx. 170g (8.6 oz)

## Contact Info:

Toll Free: 800-572-1028 (US Only)  
Phone: 714-221-9330  
Fax: 714-921-9849  
22820 Savi Ranch Parkway  
Yorba Linda, CA 92887-4610 USA  
[www.globalspecialties.com](http://www.globalspecialties.com)

Specifications and appearance subject to change without notice  
DGNV10200 052919

2 - [www.globalspecialties.com](http://www.globalspecialties.com)



Copyright © 2019, Cal Test Electronics, Inc.  
All rights reserved. Information in this publication supersedes that in all previously published material. Trade names referenced are the service marks, trademarks or registered trademarks of their respective companies.

